

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

AC/DC Adaptor

Name and address of the applicant
Nom et adresse du demandeur

BRIDGEPOWER CORP
(GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL
GWONSEON-GU SUWON-SI GYEONGGI 441-813
KOREA

Name and address of the manufacturer
Nom et adresse du fabricant

BRIDGEPOWER CORP
(GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL
GWONSEON-GU SUWON-SI GYEONGGI 441-813
KOREA

Name and address of the factory
Nom et adresse de l'usine

BRIDGEPOWER CORP
(GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL
GWONSEON-GU SUWON-SI GYEONGGI 441-813
KOREA

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Additional Information on page 2
See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

None
CTF Stage 1

Model / Type Ref.
Ref. De type

(1)ENB1010(2)(3)(4)(5)(6), (1)ENB1011(2)(3)(4)(5)(6),
BP(1)010(2)(3)(5)(6), BP(1)011(2)(3)(5)(6)
See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{eme} page

Class II
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1,
IEC 60601-1-6(ed.3), IEC 60601-1-6(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E302267-D1026-1/A0/C0-CB issued on 2016-12-19

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2016-12-23

Signature:



Jan-Erik Storgaard

For full legal entity names see www.ul.com/ncbnames

Model Details:

(1)ENB1010(2)(3)(4)(5)(6), (1)ENB1011(2)(3)(4)(5)(6), BP(1)010(2)(3)(5)(6), BP(1)011(2)(3)(5)(6)
 (Where (1), (2), (4), (5), (6) may alphanumeric, "For marketing purpose and no impact safety related critical components and constructions", where (3) may any number 05 through 48) See Test Report for details.

Factories:

WENDENG JEIL ELECTRONICS CO LTD
 XIAMEN ROAD NO.2 WENDENG ECONOMIC DEVELOPMENT ZONE WEIHAI CITY SHANDONG PROVINCE
 CHINA

BRIDGEPOWER VINA CO LTD
 LOT B9 THUY VAN INDUSTRIAL ZONE VIET TRI CITY PHU THO PROVINCE
 VIETNAM

Ratings:

Rated Input: 100-240 Vac, 50-60 Hz, 0.3 A (0.3A - 0.1A)

Rated Output:

5 Vdc, 2.0 A or 6 Vdc, 2.0 A or
 7.5 Vdc, 1.6 A or 9 Vdc, 1.1 A or
 12 Vdc, 1.0 A or 15 Vdc, 0.8 A or
 16 Vdc, 0.75 A or 18 Vdc, 0.67 A or
 24 Vdc, 0.5 A or 48 Vdc, 0.25 A

Additional Information:

Additionally evaluated to IEC 62366 (ed.1) + IEC 62366 (ed.1) am1 and EN 60601-1:2006/ A1:2013/ A12:2014.
 National Differences specified in the CB Test Report.

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-12-23

Signature:

Jan-Erik Storgaard

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

AC/DC Adaptor

Name and address of the applicant
Nom et adresse du demandeur

BRIDGEPOWER CORP
(GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL
GWONSEON-GU SUWON-SI GYEONGGI 441-813
KOREA

Name and address of the manufacturer
Nom et adresse du fabricant

BRIDGEPOWER CORP
(GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL
GWONSEON-GU SUWON-SI GYEONGGI 441-813
KOREA

Name and address of the factory
Nom et adresse de l'usine

BRIDGEPOWER CORP
(GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL
GWONSEON-GU SUWON-SI GYEONGGI 441-813
KOREA

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Additional Information on page 2
See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

None
CTF Stage 1

Model / Type Ref.
Ref. De type

(1)ENB1011(2)(3)(4)(5)(6), BP(1)011(2)(3)(5)(6)
See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

Class I
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1,
IEC 60601-1-6(ed.3), IEC 60601-1-6(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E302267-D1026-1/A0/C0-CB issued on 2016-12-19

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2016-12-23

Signature:

Jan-Erik Storgaard

For full legal entity names see www.ul.com/ncbnames

Model Details:

(1)ENB1011(2)(3)(4)(5)(6), BP(1)011(2)(3)(5)(6)

(Where (1), (2), (4), (5), (6) may alphanumeric, "For marketing purpose and no impact safety related critical components and constructions", where (3) may any number 05 through 48) See Test Report for details.

Factories:

WENDENG JEIL ELECTRONICS CO LTD

XIAMEN ROAD NO.2 WENDENG ECONOMIC DEVELOPMENT ZONE WEIHAI CITY SHANDONG PROVINCE
CHINA

BRIDGEPOWER VINA CO LTD

LOT B9 THUY VAN INDUSTRIAL ZONE VIET TRI CITY PHU THO PROVINCE
VIETNAM

Ratings:

Rated Input: 100-240 Vac, 50-60 Hz, 0.3 A (0.3A - 0.1A)

Rated Output:

5 Vdc, 2.0 A or 6 Vdc, 2.0 A or

7.5 Vdc, 1.6 A or 9 Vdc, 1.1 A or

12 Vdc, 1.0 A or 15 Vdc, 0.8 A or

16 Vdc, 0.75 A or 18 Vdc, 0.67 A or

24 Vdc, 0.5 A or 48 Vdc, 0.25 A

Additional Information:

Additionally evaluated to IEC 62366 (ed.1) + IEC 62366 (ed.1) am1 and EN 60601-1:2006/ A1:2013/ A12:2014.

National Differences specified in the CB Test Report.

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-12-23

Signature:

Jan-Erik Storgaard